

## **BLANK PAGE**






Orthopaedic Instruments and Accessories Sectional Committee, CPDC 24 [Ref: Doc: CPDC 24 (377)]

## Indian Standard

## SPECIFICATION FOR GUIDE PIN FOR KUNTSCHER NAIL

- 1. Scope Dimensions and requirements for guide pin for Kuntscher nail.
- 2. Shape and Dimensions As shown in Fig. 1.



All dimensions in millimetres.
FIG. I GUIDE PIN FOR KUNTSCHER NAIL


3. Material — Stainless steel conforming to Designation 30Cr13 of Schedule V of 1S: 1570-1961\* or the stainless steel of the following composition:

	Percent
Carbon	0·35 to 0·45
Silicon	0·6 Max
Manganese	0∙6 <i>Max</i>
Chromium	12 to 14
Nickel	I∙6 Max
Sulphur	0∙03 <i>Max</i>
Phosphorus	0·03 <i>Max</i>

- 4. Workmanship and Finish One side of the pin shall have a sharp point and on the other side a handle of suitable diameter shall be fitted tight. The pin shall be free from cracks, draw marks, pits, burrs, surface contaminations and impregnated foreign particles. It shall be passivated and polished bright to give a mirror or satin finish.
- 5. Corrosion Resistance Test Shall be according to 6.1 of IS: 5347-1969†.
- 6. Marking -- Each pin shall be clearly and indelibly marked with the following:
  - a) The name or trade-mark of the manufacturer,
  - b) 'SS' for stainless steel, and
  - c) The size of the pin.
- 6.1 ISI Certification Marking Details available from the Indian Standards Institution, New Delhi 1.
- 7. Packing Shall be as agreed to between the purchaser and the supplier.

Adopted 31 October 1969

© June 1970



<sup>\*</sup>Schedules for wrought steels for general engineering purposes. †General requirements of metal surgical implants.